

CERTIFICATE OF COMPLIANCE

Certificate Number 20180330-E494837
Report Reference E494837-20180329
Issue Date 2018-MARCH-30

Issued to: PRECISION DIGITAL CORP
233 South St
Hopkinton MA 01748

**This is to certify that
representative samples of**

PROCESS CONTROL EQUIPMENT FOR USE IN
HAZARDOUS LOCATIONS; PROCESS CONTROL
EQUIPMENT FOR USE IN ZONE CLASSIFIED
HAZARDOUS LOCATIONS
See Addendum.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: See Addendum.

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE

Certificate Number 20180330-E494837
Report Reference E494837-20180329
Issue Date 2018-MARCH-30

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Indicator Model PD66 followed by 06, 07, 08, 26, or 28, followed by L, followed by N, 2, 3 or 5, followed by N or H, followed by an additional 6-digit combination of alphanumeric and dash characters, in the form “-XXXXXX” to indicate cosmetic variations not affecting safety. For use in Class I, Division 2, Groups A, B, C, and D Hazardous or Intrinsically safe for use in Class I, Division 1, Groups A, B, C, and D when installed in accordance with manufacturer’s control drawing no. DW2516.

Class I, Zone 0, Zone 1, and Zone 2 AEx ia IIC T4 (US), Ex ia IIC T4 (Canada), Hazardous Locations Indicator Model PD66 followed by 06, 07, 08, 26, or 28, followed by L, followed by N, 2, 3 or 5, followed by N or H, followed by an additional 6-digit combination of alphanumeric and dash characters, in the form “-XXXXXX” to indicate cosmetic variations not affecting safety when installed in accordance with manufacturer’s control drawing no. DW2516.

Standard(s) for Safety:

UL 61010-1 and CAN/CSA-C22.2 No. 61010-1-12, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements.

UL 913, Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations.

UL 60079-0, Explosive atmospheres – Part 0: Equipment – General requirements.

UL 60079-11, Explosive Atmospheres – Part 11: Equipment Protection by Intrinsic Safety “i”.

UL-121201 and CAN/CSA C22.2 No. 213-17, Nonincendive Electrical Equipment for Use in Class I and II, Division 2 and Class III, Divisions 1 and 2 Hazardous (Classified) Locations.

CAN/CSA-C22.2 No. 60079-0:15, Explosive atmospheres — Part 0: Equipment — General requirements.

CAN/CSA-C22.2 No. 60079-11:14, Explosive atmospheres — Part 11: Equipment protection by intrinsic safety “i”.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/about/locations/>

