

EU DECLARATION OF CONFORMITY

Issued in accordance with ISO/IEC 17050-1:2004 and ATEX Directive 2014/34/EU.

We,

Precision Digital Corporation 233 South Street Hopkinton, MA 01748 USA

as the manufacturer, declare under our sole responsibility that the product(s),

Model PD6606 and PD6608 Loop-Powered Process Meters

to which this declaration relates, is in conformity with the European Union Directives shown below:

2014/35/EU Low Voltage Directive 2014/34/EU ATEX Directive 2014/30/EU EMC Directive 2011/65/EU RoHS Directive

This conformity is based on compliance with the application of harmonized or applicable technical standards and, when applicable or required, a European Union notified body certification.

Standards:

EN 55011:2016 EN 61010-1:2010 EN 60079-0:2018 EN 61326-1:2013 EN 63000:2018

The standards EN 55011:2016, EN 61010-1:2010, and EN 61326-1:2013 are no longer harmonized. The requirements of these standards have been checked against the harmonized standard EN 55011:2016+A2:2021, EN 61010-1:2010+A1:2019, and EN 61326-1:2021 and there were no major technical changes affecting the latest technical knowledge for the products listed above.

EC Type Examination Certificate: CML 17ATEX2015X

Product Markings:

Ex∕II1G

Ex ia IIC T4 Ga

Tamb = -40° C to $+70^{\circ}$ C

 $C \in$

ATEX Notified Body for EC Type Examination Certificate: CML B.V., NB 2776

Hoogoorddreef 15

1101BA, Amsterdam, Netherlands

ATEX Quality Assurance Notification No.: SIRA 10 ATEX M462

ATEX Notified Body for Quality Assurance: CSA Group Netherlands B.V., NB 2813

Utrechtseweg 310

6812 AR, Arnhem, Netherlands

Signed for and on behalf of Precision Digital Corporation:

Name: Jeffrey Peters

Company: Precision Digital Corporation

Title: President

Date: October 3, 2022 Document No: EU DoC PD6606 {100322}